


Modern instrumentation and practical application of flame atomic emission spectrometry

E.A. Zauer^{a*} 

Volgograd State Technical University, Volgograd 400005, Russia

* Corresponding author: zea1@live.ru



This paper belongs to a Regular Issue.

Abstract

The modern instrumentation for flame atomic emission spectrometry (FAES) is overviewed: the main technical (composition of the fuel gas used, dispersing element, number of analytical channels, reference channel, detecting element, sampling method) and analytical (determined elements, range of determined concentrations, limits and the accuracy of their determination, the duration of a single measurement, the required amount of the analyzed sample) characteristics of flame photometers for industrial and clinical use as well as spectrophotometers currently made by various manufacturers such as Sherwood Scientific Ltd., BWB Technologies UK Ltd., Labtron Equipment Ltd., Labnics Equipment Ltd. and JENWAY Ltd (UK); A.KRÜSS Optronic (Germany); Cole Parmer Instrument Company and Labfon Equipment Inc. (USA); Inesa Analytical Instrument Co., Ltd (China); OJSC Zagorsk Optical and Mechanical Plant, Unico-SIS LLC and VMK-Optoelectronics LLC (Russia); Manti Lab Solutions, Labtronics, Systonic, Globe Instruments, Electronics India, Lasany (India). The main areas of application of FAES are presented – bioenergy, agriculture (analysis of plants, soil extracts and fertilizers), mineral raw materials (geology), clinical medicine and pharmaceuticals, food industry, environmental control (analysis of drinking, technical and waste water), nuclear energy, metallurgy and chemical industry, as well as some features and problems associated with the preparation of samples for analysis by the FAES method. The review includes references to works on the practical application of FAES, published mainly from 1998 to 2023.

Keywords

flame
flame spectrometry
analysis
instrumentation
applications

Received: 31.01.24

Revised: 24.02.24

Accepted: 24.02.24

Available online: 04.03.24

Key findings

- Despite the advent of atomic absorption spectrometry and inductively coupled plasma, FAES as an analytical tool continues to be efficiently used for the determination of alkali and alkaline earth metals: the method is characterized by simplicity and speed of implementation, and its implementation requires simple and inexpensive equipment. The fields of application of the FAES method are very diverse: these are environmental control, food industry, agriculture, geology, medicine, pharmacology, nuclear and bioenergy, metallurgy and the chemical industry.

© 2024, the Authors. This article is published in open access under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).

1. Introduction

Despite more than a century and a half of the rapid development of inductively coupled plasma and atomic absorption spectrometry, FAES, as one of the most reliable, simple, convenient and economically available methods of analysis, continues to be used and retains its priority in the control of alkaline and alkaline earth elements, the need to assess the content of which in a

wide variety of objects is still quite high [1]. To a large extent, this is facilitated by the technical improvements introduced in the FAES in recent decades, aimed at automating the method, increasing its reliability and safety.

The purpose of this work is to review the current state of the FAES instrumentation and its main areas of application.

2. Modern instruments for FAES

To implement the method, flame photometers and spectrophotometers made by various manufacturers such as Sherwood Scientific Ltd., BWB Technologies UK Ltd., Labtron Equipment Ltd., Labnics Equipment Ltd. and JENWAY Ltd (UK); A.KRÜSS Optronic (Germany); Cole Parmer Instrument Company and Labfon Equipment Inc. (USA); Inesa Analytical Instrument Co., Ltd (China), OAO Zagorsk Optical and Mechanical Plant, OOO "UNIKO-SIS" and OOO "VMK-Optoelectronics" (Russia) were used. The main technical and analytical characteristics of the devices are given in Table 1-4.

Most devices for FAES use propane, butane, propane-butane mixture in cylinders or natural gas from the gas network as fuel gas. The only flame photometer on the market that can work with acetylene, which allows maintaining a higher flame temperature (2050 °C) and, accordingly, providing more accurate measurements of calcium, is offered by A.KRÜSS Optronic (Germany): the company recommends using propane to determine alkali metals (with butane as a possible substitute for propane), and acetylene for alkaline earth metals. The fuel gas flow rate is 0.4 l/min. Fuel gas regulators and fuel filters are required to control the pressure of the fuel supplied to the flame photometer and prevent particulate matter from entering the mixing chamber from the fuel source.

To achieve flame stability, constant sample supply and good accuracy of results, all flame photometers use dry, clean (oil-free) compressed air as an oxidizing agent, supplied at a pressure of 1 kgf/cm² by an oil-free air compressor connected to the photometer. Some photometers (eg BWB-XP from BWB Technologies UK Ltd.) have built-in compressors (or a low-noise built-in air pump) and electronic airflow control for optimal results. The air flow is 6 l/min.

Water separators are recommended to remove residual water from the compressed air source; they are especially needed at high humidity. The air humidity determines the amount of condensate and thus the size of the water separator (drain trap). Some photometers (eg BWB-XP from BWB Technologies UK Ltd.) have a full water separator detector.

The sample flow rate generally varies between 2 and 6 cm³/min. The analyzed samples should be aqueous solutions, homogeneous and not too viscous.

Modern flame photometers have up to five measuring channels, providing the possibility of simultaneous determination of up to five elements, respectively (photometers from BWB Technologies UK Ltd., A.KRÜSS Optronic and Labnics Equipment Ltd.).

A number of photometers have a reference channel that registers the emission line of the so-called internal standards: lithium (lithium standard) or cesium (cesium standard). The use of the reference channel makes it possible to

increase the stability of the photometer readings and eliminate deviations associated with the state of the flame, zero drift and dilution errors, i.e., ensure reproducible and accurate measurement results. The lithium reference channel is provided in photometers from Sherwood Scientific Ltd. (UK) and Cole Parmer Instrument Company (USA). In Sherwood Scientific Ltd. photometers, in addition to lithium, there is a model (photometer model M420Cs) with a cesium reference channel.

Photometers manufactured by most companies have a modular design. This makes it possible to equip the devices with additional automation components: automatic samplers (autosamplers) for working with large flows of analyzed samples (for example, DV-704 and DV-710 photometers are equipped with autosamplers for 12 and 20 positions, respectively, with a sample volume of 50 µl; Sherwood Scientific Ltd. completes its photometers with autosamplers for 40 tubes with a minimum sample volume of 200 µl; photometers from BWB Technologies UK Ltd. and A.KRÜSS Optronic can include autosamplers for 89 and 72 positions, respectively); built-in automatic dilution devices (so-called thinners) necessary when working with high concentrations requiring dilution (Sherwood Scientific Ltd., BWB Technologies UK Ltd. and A.KRÜSS Optronic) with a choice of dilution factor. A number of photometers provide: the possibility of pre-selection of the flame size (Labnics Equipment Ltd. and Labtron Equipment Ltd., UK); automatic filter selection (in Flame Photometer Superspec S20-F photometers from Spectrolab, England); built-in syringe holder for cleaning the nebulizer capillary from clogging with solutions with a high salt content (in BWB-XP photometers from BWB Technologies UK Ltd.).

In order to ensure safe operation, modern photometers are provided with the following devices: flame protection and automatic detection of flame failure (JENWAY Ltd.), flame failure detector and automatic shutdown (BWB Technologies UK Ltd., Globe Scientific Instruments): the automatic gas shutdown mechanism is activated if the flame spontaneously extinguishes.

Photometers are supplied with digital displays. For example, the BWB XP Performance Plus photometer is equipped with a four-line display, which allows calibration and determination of all elements without using a PC.

Photometers from a number of companies (for example, the BWB XP photometer from BWB Technologies UK Ltd.) feature the data transfer to a PC and the software for displaying and printing analysis results. For example, in devices from BWB Technologies UK Ltd. (BWB-XP) the software provides: mathematical signal processing (as a result, the need to use an internal standard is eliminated); calibration by two or more points; selection of the type of calibration dependences (linear and quadratic); conversion of concentrations (ppm, mg/l, meq/l and mmol/l); recalculation of concentration units of calibration curves.

Table 1 Flame photometers for laboratory applications. Technical characteristics.

Country	Manufacturer	Type (model) instrument	Fuel gas composition	Number of channels	Reference channel	Sample supply		Note
						Manual	Automatic (number of positions)	
Great Britain	Sherwood Scientific Ltd. [2]	M360	LPG (propane, butane or propane/butane mixture), or natural gas from the gas network	1	-	+	-	-
		M410	-“-	1	-	+	-	-
		M420	-“-	2	Li	+	-	-
		M420Cs	-“-	2	Cs, Li	+	-	-
		M425	-“-	2		+	+	-
	JENWAY Ltd. [3]	FPF-7	Propane, butane, propane-butane or natural gas	1		+	-	Add. Filters for Li, Ca and Ba
	BWB Technologies UK Ltd. [4]	BWB-XP	Propane, butane, LPG ^a Flow rate 0.4 l/min.	5		+	+(89)	Built-in: automatic. diluent and air compressor
		BWB XP Performance Plus	Propane, butane or natural gas. Up to 19 mbar. Consumption 0.4 l/min.	5		+	+(89)	Available with automatic sample diluent
	Spectrolab [5]	Flame Photometer Superspec S20-F	Propane, butane or natural gas. LPG ^a	4		+	-	Automatic LF ^b selection
	Labnics Equipment Ltd. [6]	Series NFP	100, 101	LPG ^a	2	-	+	-
102, 103, 104			-“-	3	-	+	-	-“-
105			-“-	4	-	+	-	-“-
106			-“-	5	-	+	-	-“-
Labtron Equipment Ltd. [7]	Series LFP A10, A11, A20, A21, A22, A30, A40	Propane/butane or propane/butane mixture. Air supply 6 l/min.	2	-	+	-	Flame size selection	
Germany	A.KRÜSS Optronic [8]	FP8000 FP8400, FP8500, FP8600 FP8700	Propane, butane, acetylene	5	-	+	+	-
USA	Cole Parmer Instrument Company [9]	Series ML 02655-00	Propane, butane, acetylene or propane/butane mixture; natural gas	1	-	+		Add. filter for Li
		02655-10	Propane/butane or propane/butane mixture. Air supply 6 l/min	2	Li	+	-	-
		02655-15	-“-	4	Li	+	-	-
	Labfon Equipment Inc. [10]	F-FPM106		2	-	+	-	-
India	Manti Lab Solutions [11]	MANTI MT 125	LPG ^a and dry oil-free air	2	-	-	-	Dual display; oil-free mini compressor block with pressure regulator

Table 1 Flame photometers for laboratory applications. Technical characteristics.

Country	Manufacturer	Type (model) instrument	Fuel gas composition	Number of channels	Reference channel	Sample supply		Note		
						Manual	Automatic (number of positions)			
India	Labtronics [12]	65	–“–	–	–	–	–	2.5 digit LED display		
		Series LT 671	–“–	1	–	–	–	For industrial and clinical applicable. Automatic ignition and filter selection		
	Systonic [13]	S-931	–“–	1	–	–	–	2.5 digit LED display. Automatic ignition system		
	Globe Instruments [14]	Globe Scientific Instruments 038-G	LPG ^a and dry oil-free air	–	–	–	–	–	–	
		1600-G	–“–	–	–	–	–	–	For industrial and clinical applicable. Embedded microprocessor	
Electronics India [15]	381	–“–	–	–	–	–	–	–“–		
Russia	OJSC Zagorsk Optical and Mechanical Plant [16]	FPA-2	Propane butane	–	–	+	–	Embedded microcomputer		
China	Inesa Analytical Instrument Co., Ltd [17]	Series FP 640	Propane butane	2					7" LCD touch screen. Flame size selection.	
		6410	–“–	2						
		6430	–“–	3			+	–		7" LCD touch screen. Automatic calculation of the correlation coefficient. Flame size selection. Optional built-in printer
		6431	–“–	3						
		6440	–“–	4						
		6450	–“–	5						

^a mixture of liquefied hydrocarbons;^b light filter.**Table 2** Flame photometers for laboratory applications. Analytical characteristics.

Country	Manufacturer	Type (model) instrument	Element concentration range (ppm)							Precision, %	Duration single measurements, s	
			Detection limits of elements (ppm)									
			Na	K	Li	Cs	Ca	Ba	Sr	Rb		
Great Britain	Sherwood Scientific Ltd. [2]	M360	1 0.1	0.5 0.1	1 0.1	–	20 2	300 20	–	–	2	≤20
		M410	0.5–10 0.02	0.5–20 0.02	0.5–20 0.02	20–100 0.2	10–100 0.2	1000 20	1000 20	1–100 0.1	<2	≤20
		M420	–“–	–“–	0–20.0 0.05	–	–	–	–	–	<2	≤20
		M420Cs	–“–	–“–	–“–	–	–	–	–	–	≤0.5	–
		M425	–“–	–“–	–“–	–	5.0–100 0.2	–	–	–	<2	≤20
	JENWAY Ltd. [3]	PFP-7	0.50–10 0.2	0.50–10 <0.25	–	10–200 <15	50–200 <30	–	–	–	≤60	–

Table 2 Flame photometers for laboratory applications. Analytical characteristics.

Country	Manufacturer	Type (model) instrument	Element concentration range (ppm) Detection limits of elements (ppm)							Precision, %	Duration single measurements, s	
			Na	K	Li	Cs	Ca	Ba	Sr			Rb
Great Britain	BWB Technologies UK Ltd. [4]	BWB-XP	0.05-1000 0.02	0.1-1000 0.05	-	2.5-1000 1	30-3000 10	-	-	-	-	
		BWBXP Performance Plus Calibra- tion	one point	0.1-60 0.03	0.05-100 0.02	0.05-50 0.02	-	1-100 0.03	5.0-100 -	-	-	30
			multi point	0.1-1000 0.03	0.05-1000 0.02	0.05-1000 0.02	-	1-1000 0.03	5-3000 1.6	-	-	-
	Spectrolab [5]	Flame Photometer Su- perspec S20-F		0-199.9 0.2	-		0-199.9 <10	-	-	-	-	
	Labnics Equip- ment Ltd. [6]	Series NFP	100	0-160 0.01	0-100 0.01	-	-	-	-	-	<3	8
			101	0-160 0.1	0-100 0.1	n/n ^b	-	-	-	-	-	-
			102	0-160 0.01	0-100 0.01	0-100 0.1	-	-	-	-	-	-
			103	0-160 0.01	0-100 0.01	-	-	0-1000 2	-	-	-	-
			104	0-160 0.01	0-100 0.01	-	-	-	0-3000 6	-	-	-
			105	0-160 0.01	0-100 0.01	0-100 0.1	-	0-1000 2	-	-	-	-
	Labtron Equip- ment Ltd. [7]	Series LFP	A10	0-160 0.184	0-100 0.156	-	-	-	-	-	-	-
			A11	-	-	opt. ^a	-	-	-	-	-	-
			A20	-	-	0-100 0.1	-	-	-	-	-	-
			A21	-	-	-	-	0-1000 2	-	-	-	-
			A22	-	-	-	-	-	0-3000 6	-	-	-
			A30	-	-	0-100 0.1	-	0-1000 2	-	-	-	-
A40			-	-	-	-	-	0-3000 6	-	-	-	
Germany			A.KRÜSS Optronic [8]	FP8000	8400, 8500, 8600	0.01-45000 0.01	0.02-45000 0.01	0.01-45000 0.01	-	0.5-45000 0.03	-	-
	8700	-		-	-	-	-	-	-	-	-	

Table 2 Flame photometers for laboratory applications. Analytical characteristics.

Country	Manufacturer	Type (model) instrument	Element concentration range (ppm)								Precision, %	Duration single measurements, s
			Detection limits of elements (ppm)									
			Na	K	Li	Cs	Ca	Ba	Sr	Rb		
USA	Cole Parmer Instrument Company [9]	Series ML	02655-00	0-1999 < 0.5	0-1999 <2	-	0-1999 5	-	-	-	-	-
			02655-10	1.00-9.99 0.5	1.00-9.99 2	-	-	-	-	-	-	-
			02655-15	1.00-9.99 0.5	1.00-9.99 2	-	0-1999 5	-	-	-	-	-
	Labfon Equipment Inc. [10]	F-FPM106	0-160 0.184	0-100 0.156	0-100 0.1	-	0-1000 2	-	-	-	1	<8
India	Manti Lab Solutions [11]	MANTI MT 125	-	-	-	-	-	-	-	-	<2300 ppm: ±2%; >2300 ppm: ±5%	-
	Labtronics [12]	Series LT	65	0-100 5	10-100 10	-	20-100 10	-	-	-	<40 ppm: ±2%; >40 ppm: ±5%	-
			671	0-100 5	10-100 10	-	20-100 10	-	-	-	-	-
	Systonic [13]	S-931	0-100 5	10-100 10	-	15-100 10	-	-	-	±2%	-	
	Globe Instruments [14]	Globe Scientific Instruments038-G	0-100 5	10-100 10	-	15-100 10	-	-	-	±2%	-	
		1600-G	0-100 0.5	-	0-100 15	50-1000 50	-	-	-	<2%	-	
	Electronics India [15]	381	0-100 5	10-100 10	-	15-100 10	-	-	-	< 40ppm ±2%; >40 ppm ±5%	-	
Russia	OJSC Zagorsk Optical and Mechanical Plant [16]	FPA-2	0.5-23 -	0.2-40 -	0.1-4.0 -	-	0.5-40 -	-	-	-	-	-
China	Inesa Analytical Instrument Co., Ltd [17]	Series FP	640	0-160 0.01	0-100 0.01	-	-	-	-	-	3	<8
			6410	0-160 0.01	0-100 0.01	-	-	-	-	-	-	-
			6430	0-160 0.01	0-100 0.01	0-100 0.1	-	-	-	-	-	-
			6431	0-160 0.01	0-100 0.01	-	-	0-1000 2	-	-	-	-
			6440	0-160 0.01	0-100 0.01	0-100 0.1	-	0-1000 2	-	-	-	-
			6450	0-160 0.01	0-100 0.01	0-100 0.1	-	0-1000 2	0-3000 6	-	-	-

^aoptional;^bnot necessary.

Table 3 Flame photometers for clinical use. Technical and analytical characteristics.

Country	Manufacturer	Type (model) instrument	Fuel gas composition	Number of channels	Sample supply		Range of determined concentrations (ppm)					Note	
					Manual	Automatic (number of positions)	Limits of detection (ppm)						
							Na	K	Li	Ca	Ba		
Great Britain	Sherwood Scientific Ltd. [2]	M410C	LPG ^a (propane, butane or propane/butane mixture), or natural gas from the gas network	1			Urine 0-4600	Urine 0-4600	0-70	-	-	-	
							Serum 2530-3910	Serum 0-230	-	-	-	-	
		M360C	-“-	1			0-1 0.1	0-0.5 0.1	0-1 0.1	0-20 2	0-300 20	-	
		M420C	-“-	2			0.5-20 0.02	0.0-20 0.05	-	-	-	-	
	JENWAY Ltd. [3]	PFP7/C	Propane, butane, propane-butane (LPG ^a) or natural gas	1	+	-	0.50-10.0 <0.2	0.50-10.0 <0.25	-	-	-	-	
Italy	Lab Service sas [18]	DIGIFLAME COMPACT Model DV-704	LPG ^a , propane or butane Compressed air: 14 l/min	3	-	+	(20)	Serum 0-230 -	Urine (0-4600) 115	0-230 69	-	-	Microprocessor. Internal valve stops gas exit when air pump is off Air pump is automatically arrested in case of lack of flame
		DIGIFLAME 2000 Model DV 710	-“-	3	-	+	(12)	0-5750 -	Urine (0-4600) -	0-230 -	-	-	-“-
India	Manti Lab Solutions [11]	MT-126	LPG ^a and dry oil-free air	2	+	-	0-100 5	10-100 10	15-100 10	-	-	Ca, Li, opt. ^b	
	Labtronics [12]	LT-66	-“-		+	-	0-4600 -	0-2300 -	0-230 -	0-46 -	-	Dual 2.5 digit LED display. Ca and Li opt. ^b	
	Systonic [13]	S-932	-“-	1	+	-	0-100 0.5	0-100 0.5	0-100 15	0-100 15	-	Dual 2.5 digit LED display. Automatic ignition system. Ca, Li, opt. ^b	
		S-935	-“-	4	+	-	0-100 0.5	0-100 50	0-100 15	0-100 15	-	Embedded software. Automatic filter selection. USB port. Ca, Li, opt. ^b	
Globe Instruments [14]	Globe Scientific Instruments 039-G	LPG ^a and dry oil-free air	2	+	-	0-100 5	10-100 10	15-100 10	-	-	Dual display Ca, Li, opt. ^b		

Table 3 Flame photometers for clinical use. Technical and analytical characteristics.

Country	Manufacturer	Type (model) instrument	Fuel gas composition	Number of channels	Sample supply		Range of determined concentrations (ppm)					Note							
					Manual	Automatic (number of positions)	Limits of detection (ppm)												
							Na	K	Li	Ca	Ba								
India	Electronics India [15]	1385	–“–	–	–	–	–	–	–	0–100	0–100	0–100	Built-in microprocessor Ca, Li, Ba						
		1382								0.5	15	50							
		1381								–	–	–	–	–	–	–	–	–	For industrial and clinical applications
		391								2	–	–	–	–	–	–	–	–	
Lasany [19]	Model 1307	–	–	–	–	–	–	–	–	–	–	–							

^amixture of liquefied hydrocarbons;^boptional.**Table 4** Flame spectrophotometers. Technical and analytical characteristics.

Country	Manufacturer	Type (model) instrument	Fuel gas composition	Operating wavelength range, nm	Analyzed elements and ranges of determined concentrations (ppm)					Number of simultaneously controlled elements	Precision, %	Single measurement duration, s	Sample consumption of the analyzed sample, cm ³	Notes
					Na	K	Li	Ca	Sr					
Russia	OJSC Zagorsk Optical and Mechanical Plant [16]	FPA-2-01	Propane butane Pressure created by the compressor, atm, no more than 0.6–1.5	580–780	0.5–23	0.5–40	0.1–4.0	0.2–40	2.5–500	1; 2; 3; 4	≤ 2.5	–	≤ 2.5	Embedded microcomputer
	OOO „UNIKO-SIS” [20]	PFA-378	Propane butane	–	–	0.5–100.0	–	15–100	–	–	±(0,036C+0,004) (C – measurement result, mg/dm ³)	≤5	≤2.5	Ba, Sr, Rb, Cs – optional
	VMK-Optoelectronics LLC «Pavlin» [21–23]	Flame spectrometer «Pavlin»	Acetylene-air Acetylene consumption 0.3–1 l/min	390–860	–	–	0.001–10 ⁵ mg/l	–	–	–	–	–	–	0.5–1.5 ml/min

JENWAY Ltd (PFP7/C photometer) and Sherwood Scientific Ltd. (photometers models M410C, M360C and M420C), in contrast to industrial (laboratory) photometers are produced for clinical use (Table 3): the display of these photometers is calibrated directly in units of Na and K concentration in biological samples (urine and blood serum); only one external standard is needed for this. Firms Globe Scientific Instruments and Labtronics (India) produce photometers designed both for laboratory and clinical use (Table 1, 3).

A number of firms (Electronics India, Manti Lab Solutions, Labtronics, Systonic, Lasany) specialize only in the production of photometers for clinical purposes (Table 3).

Flame spectrophotometers (Table 4) use a propane-butane mixture as a fuel gas. The decomposition of radiation into a spectrum and its registration are carried out using diffraction gratings and a photodiode ruler (out of 512 photodetectors), respectively. The latter provides simultaneous determination of the group of elements Na, K, Li, Ca, Ba, Ce, Rb, Sr. The processing of current information and control of the operation of photometers with the help of built-in microcomputers are provided.

The Pavlin flame spectrometer manufactured by VMK-Optoelectronics LLC (Novosibirsk, Russia) uses an air-acetylene flame, a three-slit torch made of chemically resistant stainless steel, and a stainless steel concentric atomizer similar in design to a Meinhard atomizer; the spectrum registration process is controlled using the Atom program [21–24].

3. Practical applications of FAES

A real publication boom associated with the use of flame photometry was observed in the forties and eighties of the last century. However, at present, the need for FAES is quite high.

Below is a review of the areas of application of FAES most frequently mentioned in the modern literature with a brief description of the features of sample preparation, a stage that plays an important role in analysis in general, and in analysis by the FAES method, in particular. The main emphasis is placed on works published in the last two decades.

3.1. Application in geology

Mineral raw materials are considered one of the most difficult objects for the simultaneous determination of alkali metals [25] and this is due to the fact that their content in natural samples differs by thousands of times [26]. The process of preparing samples of mineral raw materials for analysis is also time-consuming. Most often, acid decomposition with hydrofluoric acid is used for these purposes, and the resulting fluoride complexes are destroyed by evaporation, for example, with perchloric [26–30] or hydrochloric acid [31]. In [32], to determine potassium by FAES, samples of silicate rocks (basalt, monzogranite, and slate fines) are dissolved using an ultrasonic bath in a mixture of HNO₃, HCl, and HF acids.

3.2. Applications in clinical chemistry and pharmaceuticals

3.2.1. In clinical chemistry

The flame photometry method, due to its ease of implementation, rapidity and availability of equipment, is widely used in clinical laboratories to determine in the blood: sodium and potassium, the concentration ratio of which controls the functioning of muscles, including the heart [33–38]; total calcium is one of the numerous indicators of the state of human blood [39].

The FAES method has also shown its suitability for determining the concentration of Li administered for therapeutic purposes in biological fluids, for example, in bipolar disorders [40, 41]; sodium concentration and osmolality in plasma and cerebrospinal fluid (the so-called water-salt homeostasis of the brain) [42]; concentrations of sodium and potassium ions to study the physiological state of the kidneys [43]; levels of Ca, K, and Na in oral fluid [44], potassium levels in the vitreous body of the eye [45]; metals in hair [46]; calcium in urine [47], sodium and potassium in blood serum and urine [48, 49]; level of magnesium in blood plasma in patients with chronic alcoholism during withdrawal syndrome [50].

FAES method determines the intracellular content of Na⁺ and K⁺ cations to assess the functional state of erythrocytes [51].

Flame photometry examines the sodium content in dialysates and selects sodium dialysate for patients on hemodialysis [52–57]. Using flame emission photometry calibration of brain samples with a micropunch, quantitative [Na(+)] and [K(+)] maps of the brain are obtained [58].

3.1.2. In pharmaceuticals

FAES is also used in pharmaceuticals to control the sodium content in parenteral solutions used in rehydration therapy [59], sodium and potassium in powders for oral rehydration therapy [60]; potassium in complex solutions administered intravenously to replenish its intracellular losses [52], sodium in sodium diclofenac [61]; sodium and potassium in infusion solutions, such as, for example, sodium chloride solution, Ringer's solution and Reamberin [62, 63]; sodium in albumin [64].

3.3. Application in food analysis

Demand for the use of flame photometry for food analysis has grown significantly in recent years.

Since food products are complex multicomponent systems, the procedure of sample preparation is of great importance in their analysis. For liquid samples (e.g. water, beverages) direct determinations are possible with minimal sample processing such as dilution, degassing or evaporation of matrix components. Solid samples are decomposed by dry or wet ashing [65–69] as well as microwave decomposition [70].

As for the actual analysis of food products, the FAES method is known to be used to determine Li⁺, Na⁺, K⁺ and Ca²⁺ in soy sauce [71]; sodium in processed foods and bread

to assess their salinity [72–74]; Na and K in beer using microwave processing of samples [75]; calcium, magnesium and potassium in canned tomatoes [76]; rubidium in some drinks (beer, wine, vegetable and fruit juices) using a platinum wire loop for spraying in a methane-air flame [77]; sodium, potassium, lithium and rubidium in honey [78]; in the analysis of the main elements of the mineral composition of wine products [79].

3.4. Applications in the analysis of drinking, industrial and waste water

Flame photometry methods are successfully used to determine cesium and rubidium in mineral and well water using a methane-air flame [80, 81]; sodium, potassium, calcium and magnesium in natural waters [82, 83]; sodium and potassium in samples of surface, underground and river waters [84]; sodium in sea water [85]; lithium in wastewater from the production of metallic lithium and its compounds [23].

3.5. Application in bioenergy

Demand for renewable fuels has been on the rise in recent decades as fossil fuel sources are limited and, in addition, burning biodiesel results in lower emissions of carbon monoxide and hydrocarbons than burning petroleum fuels.

As with any fuel, the determination of metals in biodiesel is important because they can contribute to fuel degradation or corrosion of engine parts, reducing engine performance and life. Of particular importance is the monitoring of Na, K, Ca and Mg, whose salts and hydroxides (or their corresponding alkoxides) are used as catalysts in the biodiesel production process. These elements may be present in the form of abrasive solids or soluble soaps and contribute to engine wear due to deposit formation. In this regard, restrictions on the maximum levels of metals allowed in biodiesel have been established: for example, the maximum combined concentration of Na and K is 5 mg/kg [86].

The analysis of biodiesel is not a trivial task: its characteristics such as viscosity (11–17 times higher than that of diesel fuel), immiscibility with aqueous solutions, and high carbon content can affect the sensitivity and accuracy of the determination. FAES successfully copes with the task of monitoring the content of Na, K, Ca and Mg in biodiesel fuel.

The analysis of biodiesel by the FAES method using various types of sample preparation is widely presented in the literature: acid decomposition using HNO_3 , H_2SO_4 , HCl, and H_2O_2 , as well as their mixtures with simple or microwave [87, 88] heating; dry decomposition, which makes it possible to preconcentrate the analyte, reduce the volume of strong acids at the stage of dissolution, exclude the use of organic solvents, and shorten the stages of sample preparation, thereby contributing to a decrease in the error [89]; rapid, easy-to-implement sample dissolution in organic solvents such as ethyl alcohol [90, 91] and methyl oleate [92], which replaced xylene or n-hexane and made it possible to use aqueous spectroscopic standards that improved the

spraying process; dissolution by emulsification of a biodiesel sample with HNO_3 , n-butanol (as a co-solvent) and an aqueous solution of a surfactant (Triton X-100) [93] or by microemulsification with n-propanol and an aqueous acid solution [94]; as well as a method based on reverse-phase dispersive liquid microextraction [95].

The currently used methods for preparing biodiesel samples for elemental analysis (including the FAES method) are considered in detail in [96, 97].

3.6. Application in agriculture

The analysis of agricultural materials – plants, soil extracts and fertilizers – is another area in which FAES has been successfully applied. Traditionally, in order to determine the optimal amount of fertilizers and develop recommendations for their application, analysis of the soil and nutrient content in the foliage of plants is carried out.

3.6.1. Soil and plant analysis

Potassium and sodium are the most important elements for the vegetative growth of plants, so the control of their content in the soil and in crops is very important.

To determine the content of nutrients using the FAES method, soil samples are treated with ammonium acetate [98–103], and plant samples are treated with acids (either one (HClO_4), or a mixture of two ($\text{HNO}_3 + \text{HClO}_4$) or three ($\text{HNO}_3 + \text{H}_2\text{SO}_4 + \text{HClO}_4$) acids) [104], ultrasonic radiation [105] or subjected to extraction with a mixture of solutions of ammonium acetate – magnesium acetate [106], HCl solution [106–108], or water (since potassium in plant tissues is not bound to organic compounds and is in a soluble form) [109, 110]. Issues related to the determination of analytical elements in soil hydrolysates, soil extracts and soil solutions are discussed in sufficient detail in the book [111].

3.6.2. Fertilizer analysis

In modern agriculture, a wide range of potash fertilizers is used. FAES is employed to determine the content of potassium oxide K_2O in the production of complex fertilizers [112, 113] and organic fertilizers [114], using extraction with an ammonium oxalate solution and treatment with a mixture of nitric and perchloric acids, respectively, to decompose the samples.

3.7. Application in nuclear power, metallurgy and chemical industry

FAES has found application in the nuclear power industry for the determination of potassium in sodium used as a coolant in a fast neutron reactor [115], and alkali metals and calcium in chemical concentrates using an air-acetylene flame [24], in the chemical industry for the analysis of trace amounts of potassium in some oil samples after selective preconcentration by a centrifuge-free method of dispersive liquid microextraction, based on the use of a new target magnetic polymeric ionic liquid as a chelating and extraction solvent [116], and in metallurgy for the determination of lithium, potassium, sodium and calcium in solutions of

technological processes for the production of metallic lithium and its salts [22, 23], for determination of lithium in lithium-born alloy in the range from 59.0 to 96.0% with a relative total error of no more than 2.7% [117], and measuring the amount of rhenium in highly reducing acetylene-oxygen and acetylene-dinitrooxide flames [118].

4. Limitations

The presented review does not consider problems associated with the preparation of samples for analysis, the influence of the matrix and third elements on the results of analysis and methods for reducing this influence.

5. Conclusions

Despite the advent of atomic absorption spectrometry and inductively coupled plasma, flame emission as an analytical tool continues to be efficiently used for the determination of alkali and alkaline earth metals: the method is characterized by simplicity and high throughput (for example, the A.KRÜSS Optronic flame photometer allows up to 300 measurements per hour), its implementation requires relatively simple and inexpensive equipment. The fields of application of the FAES method are very diverse: these are environmental control, food industry, agriculture, geology, medicine, pharmacology, nuclear and bioenergy, metallurgy and the chemical industry.

In many ways, the successful use of FAES is facilitated by extensive automation of analysis, involving the automatic samplers (autosamplers) for 12–89 samples for working with large flows of analyzed samples, built-in automatic dilution devices for working with high concentration solutions, built-in syringe holder for cleaning the spray capillary from clogging with solutions with a high salt content, the possibility of pre-selection of the flame size and automatic selection of filters, an automatic detection of the loss of the flame and an automatic shutdown of the gas, which is triggered if the flame goes out, the use of digital displays that allow calibration and determination of all elements without using a PC, the ability to transfer data to a PC and print the results of the analysis, and the possibility to define up to five elements at the same time.

● Supplementary materials

No supplementary materials are available.

● Funding

This research had no external funding.

● Acknowledgments

None.

● Author contributions

The Author prepared this review by herself, including the conceptualization, data curation, formal analysis, investigation, writing, editing, and finishing.

● Conflict of interest

The authors declare no conflict of interest.

● Additional information

Author ID:

E.A. Zauer, Scopus ID [7801409703](https://orcid.org/0000-0001-9140-9703).

Website:

Volgograd State Technical University,
<https://www.vstu.ru/eng/>.

References

1. Fernández-Sánchez ML, Fernández-Arguelles MT, Costa-Fernández JM. Optical atomic emission spectrometry/flame photometry. Encyclopedia of Analytical Science (Third Edition) 2019, pp. 160–1686. doi:10.1016/B978-0-12-409547-2.14533-0
2. Sherwood-scientific (2023). Available at: <https://www.sherwood-scientific.com/products/flame-photometer/>. Accessed on 23 December 2023.
3. Jenway (2023). Available at: <http://www.jenway.com/product.asp?dsl=266>. Accessed on 23 December 2023.
4. BWB (2023). Available at: <http://flamephotometer.ru/?yclid=388458225117709598>. Accessed on 23 December 2023.
5. Spectrolab (2023). Available at: <http://www.ic-labs.com/pdfs/Other%20products/PDF/Flame%20Photometer.pdf>; <https://laboratorytalk.com/directory/12661/spectrolab>. Accessed on 23 December 2023.
6. Labnics (2023). Available at: <https://www.labnics.com/Analytical-Instruments/spectrometer/flame-photometer/nfp-100>. Accessed on 23 December 2023.
7. Labtron (2023). Available at: https://www.labtron.com/search?product_name=LFP; Flame-Photometer.pdf (yandex.ru); <https://www.labtron.com/catalog/Flame-Photometer.pdf>; <https://www.labtron.com/catalog/Flame-Photometer.pdf>. Accessed on 23 December 2023.
8. Kruss (2023). Available at: <https://www.kruss.com/en/produkte/flame-photometers/fp8700-automat-mit-verduennung/>. Accessed on 23 December 2023.
9. Cole-parmer (2023). Available at: <https://www.coleparmer.com/i/cole-parmer-dual-channel-flame-photometer/0265510>; <http://www.coleparmer.com/c/flame-photometers?searchterm=Flame%20Photometers>. Accessed on 23 December 2023.
10. Labfon Equipment Inc. (2023). Available at: https://www.labfon.com/view_catalog/F-FPM106. Accessed on 13 December 2023.
11. Manti Lab Solutions (2023). Available at: <https://www.industrybuying.com/flame-photometer-manti-lab-LA.FL8.489193>. Accessed on 23 December 2023.
12. Labtronics (2023). Available at: <https://www.indiamart.com>. Accessed on 23 December 2023.
13. Systonic (2023). Available at: <https://www.systonic.in/product/digital-flame-photometer-s-931/>. Accessed on 13 June 2023.

14. Globescientificinstruments (2023). Available at: <https://www.indiamart.com/globescientificinstruments/flamephotometers.html/>. Accessed on 13 June 2023.
15. Electronics India (2023). Available at: https://electronicsindia.co.in/products/flame_photometer/. Accessed on 13 June 2023.
16. Nd-gsi (2023). Available at: <http://nd-gsi.ru/grsi/110xx/11777-05.pdf>. Accessed on 23 March 2023.
17. Inesa analytical instrument Co., Ltd (2023). Available at: <https://inesaanalyticalinstrument.tradeindia.com/fp6450-multi-element-flame-photometer-6432271.html>. Accessed on 23 December 2023.
18. Labservicesnc (2023). Available at: <http://www.labservicesnc.com/eng/inizio-eng.html>. Accessed on 23 March 2023.
19. Lasany (2023). Available at: <https://www.indiamart.com/>. Accessed on 23 December 2023.
20. Unico-sys (2023). Available at: [http://unico-sys.ru/osnovnaya-produkciya/plamennye-fotometry/](http://unico-sys.ru/osnovnaya-produkciya/plamennye-fotometry/pfa-378/). Accessed on 13 December 2023.
21. Garanin VG, Nekljudov OA, Petrochenko DV. Programmnoe obespechenie atomno-jemissionnogo spektral'nogo analiza (programma «Atom»). *Zav Lab.* 2012;78(1):69–73. Russian.
22. Put'makov AN, Zarubin IA, Burumov ID, Seljunin DO. Spektrometr «Pavlin» dlja plamennogo atomno-jemissionnogo spektral'nogo analiza. *Zavodskaja laboratorija. Diagnostika Materialov.* 2015;81(1):105–108. Russian.
23. Matveeva AG, Gapeeva SI. Primenenie mnogokanal'nogo spektrometra «Kolibri-2» dlja analiza litievych soedinenij metodom plamennoj fotometrii. *Zavodskaja Laboratorija. Diagnostika materialov.* 2012;78(1):91–94. Russian.
24. Zarubin IA, Putmakov AN, Lukina EA, Selunin DO, Burumov ID. Extending the working range for the flame photometric determination of alkali metals and calcium using the PAVLIN spectrometer. *Analitika I kontrol [Analytics and Control]*. 2021;25(4):326–330. doi:10.15826/analitika.2021.25.4.003
25. Isaac R, Kerber JD. Atomic absorption and flame photometry: techniques and uses in soil, plant, and water analysis. (Book Chapter). *Instrumental Methods for Analysis of Soils and Plant Tissue.* 2015:17–37. doi:10.2136/1971.INSTRUMENTALMETHODS.C2
26. Shabanova EV, Zak AA, Vasil'eva IE. Probopodgotovka geologicheskikh obrazcov dlja odnovremennogo opredelenija pjati shhelochnyh jelementov metodom plamennoj atomno-jemissionnoj spektrometrii. *Zh analit himii.* 2018;73(9):671–679. Russian. doi:10.1134/S004445021809013X
27. Dolezhal Ya, Povondra P, Shul'cek Z. *Metody razlozheniya gornyh porod i mineralov /per. s chesh. k.g-m.n. Popova N.P., pod red. d.h.n. Sochevanova V.G. Moskow, Mir, 1968. 276 p. Russian.*
28. Zak AA, Shabanova EV, Vasil'eva IE. Results accuracy of the simultaneous Na, K, Li, Rb and Cs determination in geochemical objects using the Flame Atomic Emission Spectrometry. *Analitika i kontrol' [Analytics and Control]*. 2021;25(1):6–19. doi:10.15826/analitika.2021.25.1.004
29. Kurilenko LN, Kostyreva TG. Determination of quantities of sodium and potassium taking into account their mutual influence in glass and zeolites using the method of flame atomic emission spectrometry. *Glass Phys Chem.* 2016;42(3):266–269. doi:10.1134/S1087659616030068
30. Shevchenko VV, Kotsai GN. Influence of temperature on the extraction of alkali from glass powder additives to portland cement. *Glass Phys Chem.* 2017;43(5):536–537.
31. Sokolnikova JV, Vasilyeva IE, Menshikov VI. Determination of trace alkaline metals in quartz by flame atomic emission and atomic absorption spectrometry. *Spectrochim Acta Part B Atomic Spectroscopy.* 2003;58(2):387–391. doi:10.1016/S0584-8547(02)00153-2
32. Baldez DL, Avila LO, Torres DP, Martinazzo R, Silveira CAP, Vieira MA. Determination of potassium in silicate rocks by flame atomic emission spectrometry after ultrasound dissolution. *Quimica Nova.* 2018;41(10):1095–1100. doi:10.21577/0100-4042.20170290
33. Peters D, Hajes Dzh, Hift'e G. *Himicheskoe razdelenie i izmenenie. Teoriya i praktika analiticheskoy himii.* Moskow, Himiya, 1978. 816 p. Russian.
34. Buzanovskij VA. Hronologija issledovanij po metodam opredelenija koncentracii kalija v krovi cheloveka. *Obzornyj zhurnal po himii.* 2015;5(1):3 (in Russian).
35. Buzanovskij VA. Metody opredelenija kalija v krovi. *Izvestija Akademii inzhenernyh nauk im. A.M. Prohorova.* 2015;1:14–35. Russian.
36. Buzanovskij VA. Retrospektiva issledovanij po metodam opredelenija koncentracii natrija v krovi cheloveka. *Izvestija Akademii inzhenernyh nauk im. A.M. Prohorova.* 2015;2:39–65. Russian.
37. Garcia RA, Vanelli CP, Pereira Junior OdS, Corrêa JOdA. Comparative analysis for strength serum sodium and potassium in three different methods: Flame photometry, ion-selective electrode (ISE) and colorimetric enzymatic. *J Clin Lab Anal.* 2018. doi:10.1002/jcla.22594
38. El Otmani IS, Jarmoumi A, Bouatia M, Mojemmi B, Idrissi MO, Draoui M, Kamal N. Correlation study between two analytical techniques used to measure serum potassium: An automated potentiometric method and flame photometry reference method Published 2015;
39. Buzanovskii VA. Determination of calcium in blood. *Rev J Chem.* 2019;9(1):12–70. doi:10.1134/S2079978018040027
40. Mannapperuma U, Mannapperuma U, Peiris CM, Thambavita D, Galappaththy P, Pathiranaage CD, Lionel A, Jayakody RL. Validation of a flame photometric method for serum lithium estimation. *Ceylon J Med Sci.* 2017;54(2):17. doi:10.4038/cjms.v54i2.4824
41. Luzanova IS, Voznesenskaja TV, Menickaja VI, Putinskaja EV. Opredelenie sodержanija litija v bioobektah (pechen', pochki) cheloveka metodom plamennoj fotometrii. *Sudebno-medicinskaja jekspertiza.* 2007;50(5):38–39. Russian.
42. Souza LAC, Trebak F, Kumar V, Satou R, Kehoe PG, Yang W, Wharton W, Feng Earley Y. Elevated cerebrospinal fluid sodium in hypertensive human subjects with a family history of Alzheimer's disease. *Physiol Genomics.* 2020;52(3):133–142. doi:10.1152/physiolgenomics.00093.2019
43. Natochin YuV, Prokopenko AV, Kuznecova AA, Shahmatova EI. Funkcional'naya diagnostika sindroma neadekvatnoj sekrecii antidiureticheskogo gormona pri pnevmonii detej. *Pediatrics. Zhurnal im. G.N. Speranskogo.* 2020;99(2):95–101.
44. Selifonov AA, Danchuk AI. Issledovanie smeshannoj sljuny cheloveka metodom plamennoj fotometrii. *Bjulleten' medicinskih internet-konferencij.* 2015;5(12):1740.
45. Garg V, Oberoi SS, Gorea RK, Kiranjeet K. Changes in the levels of vitreous potassium with increasing time since death. *J Ind Acad Forensic Med.* 2004;26(4):136–139.
46. Bermejo-Barrera P, Moreda-Pineiro A, Moreda-Pineiro J, Bermejo-Barrera A. Acid predigestion as a slurry pretreatment for the determination of Ca, Cu, K, Mg, Na and Zn in human scalp hair by flame atomic absorption/emission spectrometry with a high-performance nebulizer. *Fresenius' J Anal Chem.* 1998;360(6):707–711. doi:10.1007/s002160050786
47. Al Omari MMH. Profiles of Drug Substances. Excipients and Related Methodology. 2016;41:31–132. doi:10.1016/bs.podrm.2015.11.003
48. Willebrands AF Jr. The determination of sodium and potassium in blood serum and urine by means of the flame photometer. 1950: doi:10.1002/recl.19500690702
49. Buzanovskij VA. Opredelenie kalija v krovi cheloveka (po materialam zhurnala CLINICAL CHEMISTRY). *Zhurnal analiticheskoy himii.* 2015;70(4):434–444. doi:10.7868/S0044450215040040

50. Stasjukinene VR, Pilvinis VK, Rejngardene DI. Gipomagnemija u bol'nyh hronicheskim alkogolizmom vo vremja abstinentsnogo sindroma. *Terapevticheskiy arhiv*. 2004;79(11):97-99. Russian.
51. Stryuk RA, Mkrtumyan AM, Bindita PL. Funkcionalnoe sostoyanie adrenoreceptorov u bol'nyh metabolicheskim sindromom. *Russ Med J*. 2008;15:1007-1012. Russian.
52. Spencer AG. Flame photometry. *The Lancet*. 1950;256(6639):623-627. doi:[10.1016/s0140-6736\(50\)91586-8](https://doi.org/10.1016/s0140-6736(50)91586-8)
53. Peitzman SJ. The flame photometer as engine of nephrology: A biography. *Am J Kidney Dis*. 2010;56(2):379-386. doi:[10.1053/j.ajkd.2010.02.343](https://doi.org/10.1053/j.ajkd.2010.02.343)
54. Turney JH, Blagg CR, Pickstone JV. Early Dialysis in Britain: Leeds and Beyond. *Am J Kidney Dis*. 2011;57(3):508-515. doi:[10.1053/j.ajkd.2010.10.043](https://doi.org/10.1053/j.ajkd.2010.10.043)
55. Ekbal NJ, Consalus A, Persaud J, Davenport A. Reliability of delivered dialysate sodium concentration. *Hemodial Int*. 2016;20:2-6. doi:[10.1111/hdi.12465](https://doi.org/10.1111/hdi.12465)
56. Shendi AM, Davenport A. The difference between delivered and prescribed dialysate sodium in haemodialysis machines. *Clin Kidney J*. 2021;14(3):863-868. doi:[10.1093/ckj/sfaa022](https://doi.org/10.1093/ckj/sfaa022)
57. Hoenig MP, Zeidel ML. Homeostasis, the Milieu Interieur, and the Wisdom of the Nephron. *Clin J Am Soc Nephrol*. 2014;9(7):1272-1281. doi:[10.2215/CJN.08860813](https://doi.org/10.2215/CJN.08860813)
58. Yushmanov VE, Kharlamov A, Yanovski B, LaVerde G, Boada FE, Jones SC. Correlated sodium and potassium imbalances within the ischemic core in experimental stroke: a ²³Na MRI and histochemical imaging study. *Brain Res*. 2013;1527:199-208. doi:[10.1016/j.brainres.2013.06.012](https://doi.org/10.1016/j.brainres.2013.06.012)
59. Pérez-López E, Alvarado PR. Implementación de un método para la determinación de sodio en soluciones parenterales por fotometría de llama (Implementation of a method for the determination of sodium in parenteral solutions by flame photometry). *Revista Tecnología en Marcha. Tecnología en Marcha*. 2017;30(4). doi:[10.18845/tm.v30i4.3414](https://doi.org/10.18845/tm.v30i4.3414)
60. Mishra P, Mahapatra MK. Comparison of sodium and potassium content in ORS powders by Flame photometric method. *Res J Pharm Biol Chem Sci*. 2011;2(3):262-267.
61. Rajendraprasad N, Basavaiah K. Sensitive Spectrophotometric and Flame Photometric Methods for Determination of Diclofenac Sodium in Pharmaceuticals. *IJAAC Int J Anal Appl Chem*. 2016;2(2).
62. Gojković V, Šalić M, Antunović V, Vučić G, Marjanović-Balaban Ž. Determination of the content of mineral substances applying different methods of chemical analysis. *Quality Life*. 2015;6(3-4):88-94. doi:[10.7251/QOL1503088G](https://doi.org/10.7251/QOL1503088G)
63. Weeks JF Jr. Flame photometric and atomic absorption determination of calcium, potassium, and sodium in Ringer's solution and injection and in lactated Ringer's solution. *J Assoc Off Anal Chem*. 1977;60(4):929-934.
64. Emel'janova IA, Kondrat'ev ML, Shkuratova OV. Validacija metoda plamennoj fotometrii dlja opredelenija natrij iona v preparate al'bumin. *Sibirskij zhurnal klinicheskoy i jeksperimental'noj mediciny*. 2011;2:91-95. Russian.
65. Carter JA, Barros AI, Nóbrega JA, Donati GL. Traditional calibration methods in atomic spectrometry and new calibration strategies for inductively coupled plasma mass spectrometry. *Front Chem*. 2018;6:504. doi:[10.3389/fchem.2018.00504](https://doi.org/10.3389/fchem.2018.00504)
66. Martínez LD, Gil RA, Pacheco PH, Cerutti S. Elemental composition analysis of food by FAES and ICP-OES (Book Chapter). *Handbook of Mineral Elements in Food*. 2015:219-238.
67. Demirel S, Tuzen M, Saracoglu S, Soylak M. Evaluation of various digestion procedures for trace element contents of some food materials. *J Hazard Mater*. 2008;152:1020-1026. doi:[10.1016/j.jhazmat.2007.07.077](https://doi.org/10.1016/j.jhazmat.2007.07.077)
68. Altundag H, Tuzen M. Comparison of dry, wet and microwave digestion methods for the multi element determination in some dried fruit samples by ICP-OES. *Food Chem Toxicol*. 2011;49:2800-2807.
69. Baker SA, Miller-Ihli NJ. Atomic Spectroscopy in Food Analysis. *Encyclopedia of Analytical Chemistry*. John Wiley & Sons, Ltd; 2006: doi:[10.1002/9780470027318.a1003](https://doi.org/10.1002/9780470027318.a1003)
70. Kingston H, Jassie LB. Introduction to Microwave Sample Preparation. ACS Press, Washington, DC, 1988.
71. Chu HT, Taylor SE. An Experimental Demonstration of a Multi-element Flame Photometer: Determination of Salt Concentration in Soy Sauce. *Int J Chem*. 2016; 8(1):25-31.
72. Chen MJ, Hsieh YT, Weng YM, Chiou RYY. Flame Photometric Determination of Salinity in Processed Foods. *Food Chem*. 2005; 91: 765-70. doi:[10.1016/j.foodchem.2004.10.002](https://doi.org/10.1016/j.foodchem.2004.10.002)
73. Vieira E, Soares ME, Ferreira IMPLVO, Pinho O. Validation of a fast sample preparation procedure for quantification of sodium in bread by flame photometry. *Food Anal Methods*. 2012;5:430-434. doi:[10.1007/s12161-011-9247-8](https://doi.org/10.1007/s12161-011-9247-8)
74. Castanheira I, Figueiredo C, André C, Coelho I, Silva AT, Santiago S, Fontes T, Mota C, Calhau MA. Sampling of bread for added sodium as determined by flame photometry. *Food Chem*. 2009;113(2):621-628. doi:[10.1016/j.foodchem.2008.07.047](https://doi.org/10.1016/j.foodchem.2008.07.047)
75. Bellido-Milla D, Moreno-Perez JM, Hernández-Artiga MP. Differentiation and classification of beers with flame atomic spectrometry and molecular absorption spectrometry and sample preparation assisted by microwaves. *Spectrochim Acta Part B Atomic Spectroscopy*. 2000; 55(7): 855-864. doi:[10.1016/S0584-8547\(00\)00164-6](https://doi.org/10.1016/S0584-8547(00)00164-6)
76. Luh BS, Niketić G. Flame photometric determination of calcium, magnesium and potassium in canned tomatoes. *J Food Sci*. 2006;24(3):305-309. doi:[10.1111/j.1365-2621.1959.tb17276.x](https://doi.org/10.1111/j.1365-2621.1959.tb17276.x)
77. Kékedy-Nagy L, Zsigmond AR, Cordoş EA. Quantification of the rubidium in beverage products micro samples by platinum-wire loop in flame atomization atomic emission spectrometry. *Acta Chim Slov*. 2010;57(4):912-915.
78. Pohl P, Stecka H, Sergiel I, Jamroz P. Different aspects of the elemental analysis of honey by flame atomic absorption and emission spectrometry: a review. *Food Anal Methods*. 2012;5(4):737-751. doi:[10.1007/s12161-011-9309-y](https://doi.org/10.1007/s12161-011-9309-y)
79. Zhiljakova TA, Dernovaja EV, Ol'hovoj JuL, Guseva IP. Primeenie atomno-absorbicionnyh i atomno-jemissionnyh metodov v analize osnovnyh jele-mentov mineral'nogo sostava vinoprodukcii. *Magarach. Vinogradarstvo i vinodelie*. 2017;3:41-43.
80. Kékedy-Nagy L, Cordoş EA. Flame atomic emission determination of rubidium in mineral and well waters using methane-air flame as excitation source. *Talanta*. 2000;52(4):645-652. doi:[10.1016/S0039-9140\(00\)00398-2](https://doi.org/10.1016/S0039-9140(00)00398-2)
81. Kékedy-Nagy L, Darvasi E. Flame atomic emission spectrometry determination of cesium in mineral and well waters using a methane-air flame. *Studia Universitatis Babeş-Bolyai. Chemia*. 2006;1:91-101.
82. Eugen D, Norbert M, Csilla S. Simultaneous determination of calcium and magnesium in natural waters by methane-air flame emission and flame atomic absorption spectrometry using a microspectrometer. *Studia Universitatis Babeş-Bolyai Chemia*. 2016;61:311-320.
83. Hajrulina AG, Temerev SV. Opredelenie natrija i kalija v prirodnyh vodah metodom fotometrii plameni. *Izvestija Altajskogo gosudarstvennogo universiteta*. 2012;3-2(75):146-149 (in Russian).
84. Banerjee P, Prasad B. Determination of concentration of total sodium and potassium in surface and ground water using a flame photometer. *Appl Water Sci*. 2020;10:113. doi:[10.1007/s13201-020-01188-1](https://doi.org/10.1007/s13201-020-01188-1)
85. Patil CS, Arbad BR. Flame photometric determination of traces of sodium content of the sea water sample. *Asian J Chem*. 2003;15(1):557-558.
86. Dancsak SE, Silva SG, Nóbrega JA, Jones BT, Donati GL. Direct determination of sodium, potassium, chromium and vanadium in biodiesel fuel by tungsten coil atomic emission spectrometry. *Anal Chim Acta*. 2014;806:85-90. doi:[10.1016/j.aca.2013.10.055](https://doi.org/10.1016/j.aca.2013.10.055)

87. Korn MGA, Santos DCMB, Guida MAB, Barbosa IS, Passos MLC, Saraiva MLMFS, Lima JLFC. Evaluation of digestion procedures for simultaneous determination of Ca, P, Mg, K and Na in biodiesel by inductively coupled plasma optical emission spectrometry. *J Braz Chem Soc.* 2010;21:2278–2284. doi:[10.1590/S0103-50532010001200015](https://doi.org/10.1590/S0103-50532010001200015)
88. Iqbal J, Carney WA, Lacaze S, Theegala CS. Metals determination in biodiesel (B100) by ICP-OES with microwave assisted acid digestion. *Open Anal Chem J.* 2010;4:18–26. doi:[10.2174/1874065001004010018](https://doi.org/10.2174/1874065001004010018)
89. de Oliveira AP, Villa RD, Antunes KCP, de Magalhães A, e Silva EC. Determination of sodium in biodiesel by flame atomic emission spectrometry using dry decomposition for the sample preparation. *Fuel.* 2009;88(4):764–766. doi:[10.1016/j.fuel.2008.10.006](https://doi.org/10.1016/j.fuel.2008.10.006)
90. Barros AI, de Oliveira AP, de Magalhães MRL, Villa RD. Determination of sodium and potassium in biodiesel by flame atomic emission spectrometry, with dissolution in ethanol as a single sample preparation step. *Fuel.* 2012;93:381–384. doi:[10.1016/j.fuel.2011.08.060](https://doi.org/10.1016/j.fuel.2011.08.060)
91. dos Santos EJ, Herrmann AB, Chaves ES, Vechiatto WWD, Schoemberger AC, Frescura VLA, Curtius AJ. Simultaneous determination of Ca, P, Mg, K and Na in biodiesel by axial view inductively coupled plasma optical emission spectrometry with internal standardization after multivariate optimization. *J Anal At Spectrom.* 2007;22:1300–1303. doi:[10.1039/b702563g](https://doi.org/10.1039/b702563g)
92. Ferreira CC, Costa LM, Barbeira P. Methyl oleate as matrix simulacrum for the simultaneous determination of metals in biodiesel samples by flame atomic emission spectroscopy. *Talanta.* 2015;138:8–14. doi:[10.1016/j.talanta.2015.02.006](https://doi.org/10.1016/j.talanta.2015.02.006)
93. Raposo JD, Costa LM, Barbeira PJS. Simultaneous determination of Na, K and Ca in biodiesel by flame atomic emission spectrometry. *J Braz Chem Soc.* 2015;26(1):147–155. doi:[10.5935/0103-5053.20140231](https://doi.org/10.5935/0103-5053.20140231)
94. Chaves ES, Saint' Pierre TD, Dos Santos EJ, Tormen L, Bascuñan VLA, Curtius AJ. Determination of Na and K in biodiesel by flame atomic emission spectrometry and microemulsion sample preparation. *J Braz Chem Soc.* 2008;19(5):856–861. doi:[10.1590/S0103-50532008000500008](https://doi.org/10.1590/S0103-50532008000500008)
95. Lourenço EC, Eyng E, Bittencourt PRS. A simple, rapid and low cost reversed-phase dispersive liquid-liquid microextraction for the determination of Na, K, Ca and Mg in biodiesel. *Talanta.* 2019;199:1–7. doi:[10.1016/j.talanta.2019.02.054](https://doi.org/10.1016/j.talanta.2019.02.054)
96. Roveda LM, Corazza M, Raposo J. Recent advances on sample preparation procedures for elemental determination in biodiesel. *Increased Biodiesel Efficiency.* 2018;127–157. doi:[10.1007/978-3-319-73552-8_6](https://doi.org/10.1007/978-3-319-73552-8_6)
97. Lepri FG, Chaves ES, Vieira MA, Ribeiro AS, Curtius AJ, De-Oliveira LCC, DeCampos RC. Determination of trace elements in vegetable oils and biodiesel by atomic spectrometric techniques – a review. *Appl Spectrosc Rev.* 2011;46:175–206. doi:[10.1080/05704928.2010.529628](https://doi.org/10.1080/05704928.2010.529628)
98. Rich CI. Elemental analysis by flame photometry. In C. A. Black et al. (ed.) *Methods of soil analysis, Part 2.* Am. Soc. of Agron., Inc., Madison, Wis. 1965;849–865. doi:[10.2134/agronmonogr9.2.c3](https://doi.org/10.2134/agronmonogr9.2.c3)
99. Hayes C. *Atomic Spectroscopy, forestry and wood products applications encyclopedia of spectroscopy and spectrometry (Third Edition).* 2017;96–104. doi:[10.1016/B978-0-12-803224-4.00153-9](https://doi.org/10.1016/B978-0-12-803224-4.00153-9)
100. Sparks DL, Page AL, Helmke PA, Loeppert RH. *Methods of soil analysis. Part 3 – Chemical methods.* Madison WI: Soil Science Society of America, American Society of Agronomy, 1996.
101. Kumar U, Mishra VN, Kumar N, Rathiya GR. *Methods of Soil Analysis.* Kalyani Publishers, Ludhiana, 2018a; 17–22.
102. Jofré FC, Perez M, Kloster N, Savio M. Analytical methods assessment for exchangeable cations analysis in soil: MIP OES appraisement. *Commun Soil Sci Plant Anal.* 2020;51(16):2205–2214. doi:[10.1080/00103624.2020.1822377](https://doi.org/10.1080/00103624.2020.1822377)
103. Jackson ML *Soil Chemical Analysis,* Prentice Hall (India), New Delhi, 1967.
104. Hafsi C, Debez A, Abdelly C. Potassium deficiency in plants: Effects and signaling cascades. *Acta Physiol Plant.* 2014;36(5):1055–1070. doi:[10.1007/s11738-014-1491-2](https://doi.org/10.1007/s11738-014-1491-2)
105. Pengo KC, Peronico VCD, De Souza LCF, Raposo JL. Feasibility of a fast and green chemistry sample preparation procedure for the determination of K and Na in renewable oilseed sources by flame atomic emission spectrometry. *Atomic Spectroscopy.* 2017;38(3):68–75. doi:[10.46770/AS.2017.03.006](https://doi.org/10.46770/AS.2017.03.006)
106. Attoe OJ. Rapid photometric determination of potassium and sodium in plant tissues. *Soil Sci Soc Am J.* 1948;12:131–134. doi:[10.2136/SSSAJ1948.036159950012000C0028X](https://doi.org/10.2136/SSSAJ1948.036159950012000C0028X)
107. Sahrawat KL. A rapid nondigestion method for determination of potassium in plant tissue. *Commun Soil Sci Plant Anal.* 1980;11(7):753–757. doi:[10.1080/00103628009367077](https://doi.org/10.1080/00103628009367077)
108. Sahrawat KL. Potassium determination in grain samples using the nondigestion (dilute HCL extraction) method. *Commun Soil Sci Plant Anal.* 1984;15(1):81–86. doi:[10.1080/00103628409367455](https://doi.org/10.1080/00103628409367455)
109. Rosolem CA, Calonego JC, Foloni JSS. Potassium leaching from millet straw as affected by rainfall and potassium rates. *Commun Soil Sci Plant Anal.* 2005;36(7):1063–1074. doi:[10.1081/CSS-200050497](https://doi.org/10.1081/CSS-200050497)
110. Reddy DD, Veeranki K. Simple and inexpensive water extraction method for assaying potassium concentration in tobacco plant tissue. *Commun Soil Sci Plant Anal.* 2013;44:962–970. doi:[10.1080/00103624.2012.747603](https://doi.org/10.1080/00103624.2012.747603)
111. Wright R.J., Stuczynski T.I. Atomic absorption and flame emission spectrometry. In book *Methods of Soil Analysis, Part 3: Chemical Methods.* 2018;65–90. doi:[10.2136/sssabookser5.3.c4](https://doi.org/10.2136/sssabookser5.3.c4)
112. Zhenghong Z. *Method for analysis of potassium oxide content in compound fertilizer technology.* Chemistry Petrochemical Industry Application. 2013.
113. Hongjun L. *Rapid analysis of K₂O content in compound fertilizers.* Chemistry Phosphate and Compound Fertilizer. 2003.
114. Zakiyah Z, Rahmawati C, Fatimah I. Analysis of phosphorus and potassium levels in organic fertilizer. In the integrated laboratory of jombang district agriculture office environmental science. *Indones J Chem Res.* 2018;3(2):38–48. doi:[10.20885/ijcr.vol3.iss2.art1](https://doi.org/10.20885/ijcr.vol3.iss2.art1)
115. Xie C, Wen X, Jia Y, Sun S. [Determination of potassium in sodium by flame atomic emission spectroscopy]. *Guang Pu Xue Yu Guang Pu Fen Xi.* 2001;21(3):366–369.
116. Beiraghi A, Shokri M. A novel task specific magnetic polymeric ionic liquid for selective preconcentration of potassium in oil samples using centrifuge-less dispersive liquid-liquid microextraction technique and its determination by flame atomic emission spectroscopy. *Talanta.* 2018;178:616–621. doi:[10.1016/j.talanta.2017.08.080](https://doi.org/10.1016/j.talanta.2017.08.080)
117. Kalinina AA, Konopkina IA, Vahnina OV, Koroleva IV, Zhogova KB, Annikova SA. Vybor metodik opredelenija litija i bora v litij-bornom splave. Zavodskaja laboratorija. Diagnostika materialov. 2023;89(1):20–27. doi:[10.26896/1028-6861-2023-89-1-20-27](https://doi.org/10.26896/1028-6861-2023-89-1-20-27)
118. Zaytseva PV, Pupyshev AA, Evdokimova OV, Shunyaev KYu. To question of the rhenium determination by flame atomic absorption and atomic emission spectrometry. *Analitika i kontrol' [Analytics and Control].* 2012;16(1):30–38. Russian.